

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

AUG 14 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130288</u> ✓	
Name <u>DIXIE SPENCER</u>		ID# <u>2280</u>	Date <u>8/12/09</u>
A	Agency <u>AST ANIAK</u>	Phone # <u>675-4598</u>	
Instrument Location <u>ANIAK</u> ✓			
Alco S/N <u>56799</u> ✓		Target Value <u>.084</u> ✓	High Pressure <u>110.0</u>
B	Alco Test Values		
	<u>.084</u> ✓ 1 <sup>st</sup> Alco	<u>.084</u> ✓ 2 <sup>nd</sup> Alco	
Signature <u>Dixie Spencer</u> ✓		<u>CS</u> <u>8/12/09</u>	
(OVER)			


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of Sept, 2009.  
Carolyn M. Noland (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130288

8/14/01  
CWA

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130288

AUGUST 12, 2000

OPERATOR'S NAME:  
SUBJECT NAME:  
TEST NO.: 2059  
DATE: 8/14/01

SUBJECT'S FIRST NAME/MI:  
SUBJECT:

ALC. #: 1234567  
DEPT/AGENCY: ANTI  
CASE REPORT: 09-122222  
TEST TYPE: T  
ALCO TARGET VALUE: .034  
ALCO S/N: 56799

--- BREATH ANALYSIS ---

.034 ADJUSTED FOR 29.77 in  
ALCO TARGET .033 09:05  
BLANK TEST .000 09:06  
INTERNAL STANDARD .033 09:06  
ALCO TO 29.77 in .034 09:06  
BLANK TEST .000 09:07  
SUBJECT SAMPLE .036 09:09  
BLANK TEST .000 09:10  
ALCO TO 29.77 in .034 09:10  
BLANK TEST .000 09:11

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130288

AUGUST 12, 2000  
TIME 09:05

--- TEST ---  
ALC. # 1234567  
ALCO TARGET .033  
BLANK TEST .000  
INTERNAL STANDARD .033  
ALCO TO 29.77 in .034  
BLANK TEST .000  
SUBJECT SAMPLE .036  
BLANK TEST .000  
ALCO TO 29.77 in .034  
BLANK TEST .000  
SOFTWARE DATE: 02/28/01  
HEATERS  
SAMPLE CHAMBER: 43c  
BREATH TUBE: 43c  
BAROMETER: 29.77 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

11/12/01 (1) 11/12/01 (1) 11/12/01 (1) 11/12/01 (1)  
ALCO TARGET .033 09:05  
BLANK TEST .000 09:06  
INTERNAL STANDARD .033 09:06  
ALCO TO 29.77 in .034 09:06  
BLANK TEST .000 09:07  
SUBJECT SAMPLE .036 09:09  
BLANK TEST .000 09:10  
ALCO TO 29.77 in .034 09:10  
BLANK TEST .000 09:11